



HAL
open science

How to enhance learning of robotic surgery gestures? A tactile cue saliency investigation for 3D hand guidance. (“Best Poster Award”)

Gustavo D. Gil, Julie M. Walker, Nabil Zemiti, Allison M. Okamura, Philippe Poignet

► To cite this version:

Gustavo D. Gil, Julie M. Walker, Nabil Zemiti, Allison M. Okamura, Philippe Poignet. How to enhance learning of robotic surgery gestures? A tactile cue saliency investigation for 3D hand guidance. (“Best Poster Award”). HSMR: Hamlyn Symposium on Medical Robotics, Jun 2019, London, United Kingdom. lirmm-02093730

HAL Id: lirmm-02093730

<https://hal-lirmm.ccsd.cnrs.fr/lirmm-02093730>

Submitted on 9 Apr 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L’archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d’enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

How to enhance learning of robotic surgery gestures? A tactile cue saliency investigation for 3D hand guidance

Gustavo D. Gil¹, Julie M. Walker², Nabil Zemiti¹, Allison M. Okamura², Philippe Poignet¹

¹ LIRMM, University of Montpellier, CNRS, Montpellier, France

² Department of Mechanical Engineering, Stanford University, Stanford, CA 94305, USA

gustavo.gil@lirmm.fr

INTRODUCTION

The current generation of surgeons requires extensive training in teleoperation to develop specific dexterous skills, which are independent of medical knowledge. Training curricula progress from manipulation tasks to simulated surgical tasks [1]–[3] but are limited in time due to the need for proper mentoring for each trainee.

We propose to integrate surgical robotic training together with Haptic Feedback (HF), as is illustrated in Fig. 1. The reason is that a good use of HF during training can improve skill acquisition [4]–[7].




Figure 1. Full dry and wet lab approach (non-clinical setup) of our complete robotic surgery training system. During the execution of the surgical gestures the highlighted haptic feedback loop transmit to the surgeon trainee timely information about his/her performance.

In this paper we introduce an exploratory work on a portable haptic device designed to offer hand guidance in 3D space. The idea is that the HF can be felt by the trainee “as if” a force was directing his/her entire hands through of the surgical gesture execution.

Mainly, we create HF signals by stretching the skin of a pair of fingers (e.g., Thumb-Index). Our haptic device acts on the *finger pads* of the *fingertips* (see Fig. 2), in order to stimulate the cutaneous mechanoreceptors.




Figure 2. The colored areas indicate the jargon for different zones of the human fingertips (i.e., finger tip and finger pad).

MATERIALS AND METHODS

The device consists of a pair of servomotors mounted in a 3D printed handle, as shown in Fig. 3. Each motor rotates a lever arm to stimulate the trainee’s finger pads of the thumb and index distal phalanges, while the trainee holds it in his/her hand.

The stimulations, called *haptic signals*, consist of to stretch the skin on the finger pads by rotating through a semicircular arc (see schema in Fig. 3). Thus, different movements of the arms give to the user the sensation that their hand is being pulled in certain directions. We called these sensations *haptic cues*.




Figure 3. Scheme for one actuator of the haptic device (working principle) and views of the entire device.

The experiment presented here aims to identify two key aspects: (1) cue saliencies or cue strength; and (2) if there is a common perception of tactile cues between users. The first aspect is related to stimulus-cue repeatability by the users, while the second one is related to the identification of a common stimulus-cue set among different users.

We investigate the feasibility to induce three types of hand movements, or 3 Degrees of Freedom (DoF). Namely: right/left wrist twisting; hand moving forward/backward; and hand tilting up/down.

The experiment involved 8 right-handed people. 24 different stimuli were applied in a pseudorandom sequence of 240 stimuli per trial (each stimulus 10 times). Trials included two time breaks in order to release the device, relax the hand, and continue the experiment.

For each *haptic signal* (i.e., stimulation induced), users reported the cue felt from the following six options: (1-2) hand moving Forward/Backward; (3-4) Tilt Up/Down; and (5-6) Twist Left/Right.

During the trial, the users trigger each haptic stimulus by using a key of a computer keyboard. Then, the cue felt was reported by a “number key” assigned.

Summarizing the test method, we executed a system identification procedure to investigate the responses of the tactile sense of each user. Our aim was to identify stimuli-cues that can be used as commands to guide the user’s hand in 6 different directions (i.e., in 3 DoF).

RESULTS

Each trial took about 45 minutes plus extra time for the familiarization with the procedure. People were constrained to select between 6 possible directions. It was noticeable that most of the time participants chose a direction after triggering each stimulus just 1 or 2 times. The multidimensional nature of the results is expressed by colored marks in a two-dimensional map (Fig. 4). The meaning of the features of this map are:

- the colors encode the type of cue most often selected by the users (e.g., hand pulled Forward/Backward, hand Tilt Up/Down, and hand Twist Left/ Right).
- the x-y location of each mark corresponds to a combination of angular displacement for each servomotor (stimulus related).
- the mark size represents the cue saliency or cue clarity. The bigger diamonds correspond to the most salient cues.




Figure 4. Cues identified by 8 users in the exploratory experiment. The mark size relates to cue saliency, the color to cue type, and the location to stimulus direction and magnitude (servomotor actuation).

From the data obtained in this exploratory study, we can agree that the device is able to induce at least two clear tactile cues, called hand Twist Left and hand Twist Right. These two cues are very salient for several combinations of servomotor values (see the 2nd and 4th quadrants of the Actuation-Cues Map), identified more than 85% of the time.

Conversely, the cues labeled as hand Tilt Up and hand moves Backward, are not as salient. For the remaining cues, hand moves Forward and hand Tilt Down, a trend appears, so it seems possible to convey this information to different users.

DISCUSSION

Interestingly, our haptic device is able to induce a directional cue (i.e., a feeling of direction) without prior training. Thus, the principle of skin stretch [8], [9] explored here allows us to envisage ungrounded and low-cost haptic devices for hand guidance.

In the experiment some people had problems with the thumb, which occasionally slipped out from the haptic device. This issue can be related to the symmetrical arrangement of the device. A modified device with a better ergonomics is foreseen for future test campaigns. A good alternative is proposed in [10].

To conclude, we see the Haptic Feedback (HF) as a key ingredient of our recipe for surgical robotic training. There is potential for HF to speed training times for surgeon residents, while increasing development of relevant visual-motor skills. Our ongoing work is related to integrate the HF in the RAVEN II platform.

REFERENCES

- [1] A. Volpe *et al.*, “Pilot Validation Study of the European Association of Urology Robotic Training Curriculum,” *Eur. Urol.*, vol. 68, no. 2, pp. 292–299, Aug. 2015.
- [2] S. V. Whitehurst *et al.*, “Comparison of Two Simulation Systems to Support Robotic-Assisted Surgical Training: A Pilot Study (Swine Model),” *J. Minim. Invasive Gynecol.*, vol. 22, no. 3, pp. 483–488, Mar. 2015.
- [3] G. Whittaker *et al.*, “Validation of the RobotiX Mentor Robotic Surgery Simulator,” *J. Endourol.*, vol. 30, no. 3, pp. 338–346, Mar. 2016.
- [4] L. M. Crespo and D. J. Reinkensmeyer, “Haptic Guidance Can Enhance Motor Learning of a Steering Task,” *J. Mot. Behav.*, vol. 40, no. 6, pp. 545–557, Nov. 2008.
- [5] M. Mulder, D. A. Abbink, and E. R. Boer, “The effect of haptic guidance on curve negotiation behavior of young, experienced drivers,” in *2008 IEEE International Conference on Systems, Man and Cybernetics*, Singapore, Singapore, 2008, pp. 804–809.
- [6] J. Bluteau, S. Coquillart, Y. Payan, and E. Gentaz, “Haptic Guidance Improves the Visuo-Manual Tracking of Trajectories,” *PLoS ONE*, vol. 3, no. 3, p. e1775, Mar. 2008.
- [7] A. Teranishi, G. Korres, W. Park, and M. Eid, “Combining Full and Partial Haptic Guidance Improves Handwriting Skills Development,” *IEEE Trans. Haptics*, vol. 11, no. 4, pp. 509–517, Oct. 2018.
- [8] H. Olausson, I. Hamadeh, P. Pakdel, and U. Norrsell, “Remarkable capacity for perception of the direction of skin pull in man,” *Brain Res.*, vol. 808, no. 1, pp. 120–123, Oct. 1998.
- [9] B. T. Gleeson, S. K. Horschel, and W. R. Provancher, “Perception of Direction for Applied Tangential Skin Displacement: Effects of Speed, Displacement, and Repetition,” *IEEE Trans. Haptics*, vol. 3, no. 3, pp. 177–188, Jul. 2010.
- [10] J. M. Walker *et al.*, “Holdable Haptic Device for 4-DOF Motion Guidance,” ArXiv [cs], 2019.