

MoS2 as the Sensing Platform for the Non-Enzymatic Detection of Cortisol: A First-Principles Study

Gabriele Boschetto, Stefania Carapezzi, Aida Todri-Sanial

▶ To cite this version:

Gabriele Boschetto, Stefania Carapezzi, Aida Todri-Sanial. MoS2 as the Sensing Platform for the Non-Enzymatic Detection of Cortisol: A First-Principles Study. ImagineNano2021, Nov 2021, Bilbao, Spain. lirmm-03430916

HAL Id: lirmm-03430916 https://hal-lirmm.ccsd.cnrs.fr/lirmm-03430916

Submitted on 16 Nov 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

MoS₂ as the Sensing Platform for the Non-Enzymatic Detection of Cortisol: A First-Principles Study

Gabriele Boschetto

Stefania Carapezzi, Aida Todri-Sanial

CNRS-LIRMM, University of Montpellier, 34095 Montpellier, France

<u>gabriele.boschetto@lirmm.fr</u>; stefania.carapezzi@lirmm.fr; aida.todri@lirmm.fr

Atomically thin two-dimensional (2D) materials have been —and are still currently being— extensively studied due to their unique mechanical, electrical, and optical properties, which, together with their ultraenable the development of thin size, compact devices and innovative technologies. Within the vast chemical space of transition metal dichalcogenides (TMDs), single-layer molybdenum disulphide (MoS₂) is with no doubt one of the most studied material due to its stability and its direct optical band gap of 1.8 eV, which make it the ideal candidate to be used in a wide range of nanoelectronic devices, going beyond conventional CMOS technology. [1] Here we look at MoS2 in the context of biosensing, and we study such material as component of core field-effect biosensors (Bio-FETs) for the detection of cortisol. Ultimately, the aim of this study is to design and integrate such biosensors in wearable health monitoring devices. [2] We want to bridge the gap between materials' properties and device physics and to do so, carry out first-principles atomistic computer simulations in the framework of density functional theory (DFT). Our study constitutes the first step of a wider multi-scale modelling approach in which the goal is to construct a full atomistic-to-device level model.

Recently, MoS₂ has been studied as a sensing platform for detecting mainly gas and small biological molecules, such as glucose. [3] Enzymatic biosensing is the most common approach, however, non-enzymatic sensing can provide higher sensor stability and

prompt response, at the expense chemical selectivity. Here, we are interested in the non-enzymatic detection of cortisol in human sweat as a mean to monitor the risk of cardiovascular diseases. However, the mechanisms that govern the interaction between the analyte and MoS2 at the molecular level are far from understood. Thus, we thoroughly explore the MoS₂/cortisol interaction in terms of both structural, electronic, and charge transfer properties viable to assess sensina mechanisms. We study the impact of some of the most used metal dopants employed in lab-scale experiments, such as Ni, Pt, Pd, in order to modulate the sensing platform with respect to bare MoS2. In addition to singleatom doping, we also explore the use of metal nanocluster (e.g., Pt and Au) to decorate the MoS2 layer as yet another mean to detect cortisol.

Overall, our work ultimately aims to obtain a deep understanding of the properties of MoS_2 when used as a sensor to drive the design of devices towards better performance.

References

- [1] Y. Qiao et al., "Fabricating molybdenum disulfide memristors," ACS Appl. Electron. Mater., 2, 346-370, **2020**.
- [2] EU H2020 SmartVista project, www.smartvista.eu.
- [3] G. Jeevanandham et al., "Nickel oxide decorated MoS₂ nanosheet-based non-enzymatic sensor for the selective detection of glucose," *RSC Adv.*, 10, 643-654, **2020**.